

COMPETITIVE PROBLEMS IN THE DRUG INDUSTRY 9083

<u>Product</u>	<u>Reason</u>
Uric acid set	Decomposition resulting in inaccurate readings
Serum iron set	Visible mold growth
Salicylic acid	Defective
Cholesterol set	Labeling error

FDA officials stated that, because the defects in the products could lead to inaccurate medical diagnoses, all district offices were requested to check on the effectiveness of these recalls.